



CEM/CMO/2020/031

PPE Dedicated Supply Channel Important Customer Alert

Non-sterile gowns Flosteril FLO-MED-8130

ICA:005

Date: 11 August 2020

Alert

Flosteril FLO-MED-8130 non-sterile gowns do not meet minimum standards for fluid resistance.

Products affected

There is one product affected. Flosteril FLO-MED-8130 Non-Sterile gowns supplied by Vannin Healthcare (VHC) Global. Packaging for this product may vary. Images of the known variations of the packaging are in Annex 1 to help identification.

Affected Product Stock Keeping Units (SKU):

BWK9877

BWK9845

BWK9846

BWK9847

BWK9823

BWK9796

BWK9741

Supplier: Vannin Healthcare (VHC) Global Flosteril

MPC: FLOMED-8130

Action

Provided from central stocks:

- Check your stock immediately for unused Flosteril FLO-MED-8130 gowns from the SKUs listed above
- Stop using immediately and quarantine any unused gowns.
- Prepare quarantined gowns for collection. Further instruction will be issued in the next two weeks for their collection with a target date of collection to be started week commencing 31 August 2020.
- Do not dispose of the gowns.
- Share this ICA with anyone you supplied these gowns to.

Privately sourced:

- Check your stock immediately as above. Quarantined stock shall be managed locally and should be returned to your supplier

Contact NSDR on **0800 915 9964** for any emergency replacements required

For all general enquiries for the PPE Dedicated Channel, call Customer Services on 0800 876 6802.

Annex 1 Packaging Images. Packaging may vary.

